

Written Quiz

	<u>POINTS</u>
1. The two types of resuscitators are _____ and _____.	<u>20</u>
2. The range of safe respirable pressures for a demand valve resuscitator is _____.	<u>20</u>
3. Masks should be transparent to visualize _____.	<u>20</u>
4. Steel and aluminum tanks need to be hydrostatically tested every _____ years.	<u>20</u>
5. In the U.S., O2 tanks are colored _____.	<u>20</u>

POINTS POSSIBLE: 100

POINTS DEDUCTED:

FINAL SCORE:

	<u>POINTS</u>
1. The two types of resuscitators are BAG VALVE MASK and DEMAND VALVE RESUSCITATORS .	<u>20</u>
2. The range of safe respirable pressures for a demand valve resuscitator is 35 - 60 CM/H2O .	<u>20</u>
3. Masks should be transparent to visualize VOMITUS/SECRETIONS .	<u>20</u>
4. Steel and aluminum tanks need to be hydrostatically tested every FIVE years.	<u>20</u>
5. In the U.S., O2 tanks are colored GREEN .	<u>20</u>

POINTS POSSIBLE: 100

POINTS DEDUCTED:

FINAL SCORE: